

A new HGA-FLVQ model for *Mycobacterium Tuberculosis* detection

N. Charibaldi,^{1*} A. Harjoko,² Azhari² and B. Hisyam³

¹Doctoral Program, Department of Computer Science and Electronics, Universitas Gadjah Mada, Yogyakarta, Indonesia.

²Department of Computer Science and Electronics, Universitas Gadjah Mada, Yogyakarta, Indonesia.

³Faculty of Medicine, Universitas Gadjah Mada, Yogyakarta, Indonesia.

*E-mail: novrido.charibaldi@mail.ugm.ac.id.

This paper was edited by Ha-Duong Ngo.

Received for publication April 30, 2018.

Abstract

This research aims to develop an MTB detection model from the FLVQ neural network to HGA-FLVQ model. In this research, the FLVQ method was developed through strengthening its initiation, in which the first cluster centers used as FLVQ input were optimized first by HGA. The results show that sensitivity and specificity of the HGA-FLVQ model reach 96.30 and 95.65%, whereas the sensitivity of an FLVQ method is 70.83%, and the sensitivity of an LVQ method is 87.50%. The specificity of an FLVQ method and the specificity of an LVQ method are 84.62%. Based on these results, we can say that the HGA-FLVQ model is better than FLVQ and LVQ methods. It also means that relative amplitude can be used by the HGA-FLVQ model as a feature to detect the presence of MTB in the sputum of TB-suspected patients. Thus, the HGA-FLVQ model can be used to strengthen TB laboratory examination at Public Health Centers in Indonesia.

Keywords

MTB detection, HGA-FLVQ model, FLVQ method, HGA, Relative amplitude.

The first generation tuberculosis (TB) laboratory examination method (Charibaldi dan and Harjoko, 2013) occasionally used in developing countries is based on Ziehl-Neelsen (ZN) staining as the gold standard, with 50 to 60% sensitivity (Fend et al., 2006). Nevertheless, in which TB negative samples examined as positive ZN staining could occur infiltration (Gibson et al., 2009). This ZN staining method works microscopically. Research (Kolk et al., 2010) concluded that the ZN staining method still has unsatisfying result; moreover, TB has become a global public health threat and a health emergency in developing countries. Hence, the development of TB laboratory examination method, which is more sensitive, quicker, more applicable, more portable, and cheaper (Bruins et al., 2013; Zhou et al., 2011), is needed.

The TB laboratory examination methods are newer compared to the ZN staining method. The first type of TB laboratory examination method is based on nucleic acid amplification, such as polymerase

chain reaction (PCR), strand displacement amplification (SDA), and transcription-mediated amplification (TMA). The second type is based on serology and cytokine released metal rate examination. Both methods are considered to be more advanced. These two methods are commonly available in developed countries, researching continually TB examination methods. In developing countries, examination methods other than ZN staining are considered as complicated and expensive to be applied widely. Hence, a new approach, therefore, is required for the development of the TB laboratory examination methods (Pavlou et al., 2004). There is a TB laboratory examination method that uses thorax X-ray irradiation (also called the generation; Charibaldi dan and Harjoko, 2013). This method has joined in the TB laboratory examination procedure in some developing countries, but this laboratory examination method is not suitable to be implemented in countries with low income yet with high prevalence.

The next generation of the development of the TB laboratory examination method has used an electronic-nose device by detecting a volatile organic compound (VOC). VOC is an organic compound evaporating quickly. A VOC is hypothesized reasonably as containing an active pulmonary TB biomarker that is derived from infected organisms (a metabolite of TB *Mycobacterium*) or from infected hosts (a result of oxidative stress) (Phillips et al., 2010). The main problem in the identification of VOC is the pattern similarity of the electronic-nose response curve resulting from the low sensitivity of the sensing system selectivity (Polikar et al., 2001).

Previous research studies aimed at the development of the method detecting the presence of *Mycobacterium tuberculosis* (MTB) in patients' sputum using an electronic-nose sensing device. These research studies were carried out using various ways to produce different specificities and sensitivities, as presented in Table 1.

The artificial neural network is implemented to support the development of the TB laboratory examination method. An artificial neural network provides a solution with strong robustness, for approximating real values, discrete values, or vector-valued goal function (Wang and Liu, 2014). An artificial neural network is so far known as the most effective way, for certain types of problems, such as learning to explain complex real-world sensor data (Wang and Liu, 2014). Learning vector quantization (LVQ) is a neural network prototype-based learning (Elly et al., 2013). The prototypes are determined in a training process from training dataset and can be interpreted in a straight-

forward way as they capture essential features of the data in the same space (Elly et al., 2013).

Therefore, it is necessary to elaborate artificial neural network potentials in some research works, especially to differentiate three different odors and to recognize mixed scents (Kusumoputro et al., 1999). The previous analysis (Kusumoputro et al., 1999) was carried out in an attempt to resolve the weaknesses of LVQ artificial neural network by using fuzzy learning vector quantization (FLVQ) artificial neural network. A result of the research showed that the odor recognition probability of FLVQ artificial neural network is quite high to distinguish the scent of oranges, roses, and cananga flowers. The study (Kusumoputro and Jatmiko, 2002) also revealed that the recognition of mixed odors on the electronic sensing system using the algorithm of FLVQ artificial neural network is better than back propagation (BP) artificial neural network and probabilistic neural network (PNN). The research (Kusumoputro et al., 2002) was carried out to observe the performance of FLVQ artificial neural network for recognizing the mixed odors. Then, the recognition rate of FLVQ artificial neural network was compared to the recognition rate of BP artificial neural network. The combined scents tested in the research consisted of two types of mixtures: first, mixture of two odors and, second, the mixture of three scents. According to this research (Kusumoputro et al., 2002), the result of the recognition of the two or even three-odor mixtures obtained using FLVQ artificial neural network is better compared to that obtained using BP artificial neural network and PNN.

Table 1. Previous researches.

Reference	Sample type	Classification methods	Results
Fend et al. (2006)	Sputum	Back propagation artificial neural network	Sensitivity 89.09% Specificity 91.14%
Gibson et al. (2009)	Sputum	Linear discriminant analysis	Average Sensitivity 80% Average Specificity 75%
Kolk et al. (2010)	Sputum	Rob electronic-nose: Linear Discriminant Analysis Partial least square discriminant analysis	Sensitivity 57–64% Specificity 61–70% Sensitivity 42–50% Specificity 73–77%
		Walter electronic-nose: Linear discriminant analysis Partial least square discriminant analysis	Sensitivity 56–66% Specificity 65–68% Sensitivity 37–61% Specificity 56–67%




Figure 1: An HGA-FLVQ Model Block Diagram.

Methods

Main points relating to the development of model are described in this section. First, the below equation based on (Dahlan, 2010) calculating the sample size for this research is explained. Second, the developed block model and tasks of each block are elaborated. Third, various processes of the overall development of model are discussed from the beginning to the end. The developed model in this research is presented in Figure 1.

In this research, the sample size was calculated based on (Dahlan, 2010) the following equation:

$$n = \frac{(Z_{\alpha})^2 Sen(1-Sen)}{d^2 P} \quad (1)$$

Here, n denotes the research sample size, Sen is the expected sensitivity of the model of which the diagnostic value was tested, d is expected research precision, α is the accepted error rate to determine the deviation standard (Z_{α}), and P is the prevalence of TB disease (the proportion of positive-TB patients among the TB suspects).

The following is the implementation of Equation (1) in this research. The error rate of type 1 (one direction) that is still accepted (α) equals 5%. The deviation standard (Z_{α}) is 1.96. The expected sensitivity is 90%. The expected research precision (d) is 10%. The TB prevalence (P), referring to the proportion of positive-TB patients after an assessment of overall TB suspects in 2015 who received medical treatments from Prof. Dr. Barmawi Hisyam, Sp.P.D.-KP, equals 70%. Hence, the research sample size is as follows:

$$\begin{aligned} n &= \frac{(1.96)^2 (0.9)(1-0.9)}{(0.1)^2 (P)} = \frac{(3.8416)(0.09)}{(0.01)P} \\ &= \frac{34.5744}{P} = \frac{34.5744}{0.7} = 49.392 \end{aligned}$$

It is rounded up to 50 TB suspects.

After receiving Ethics Committee Approval from Medical and Health Research Ethics Committee (MHREC), Faculty of Medicine, Universitas Gadjah Mada, with Ref number: KE/FK/71/EC/2016, primary

data retrieval at this research could be started in TB laboratory, Faculty of Medicine, Universitas Gadjah Mada.

The e-nose device employed in this research consisted of six sensors, namely, TGS822, TGS813, TGS2611, TGS825, TGS826, and TGS2620. However, primary data (training data and testing data) of this research were only obtained from three sensor response curves. The sensors used in this research were TGS822, TGS813, and TGS2611, as the other sensors were not adequately responsive and were selective to significantly distinguish the characteristics of patients (relative amplitudes) of the positive ZN staining and the negative ZN staining, using direct eye observation. Each period of the electronic-nose response curve of the three sensors was sampled every 5 sec.

There were three categories of samples, namely 49 negative ZN staining subjects' sputum samples, 49 positive ZN staining patients' sputum samples, and 50 sputum samples of TB-suspected patients who received medical treatments at Public Health Centers in Yogyakarta and A RESPIRA Pulmonary Hospital of Yogyakarta Province between February 25 and May 8, 2016. An inclusion criterion of negative-TB samples was when a patient was diagnosed as a TB suspect by the doctor, but the patient was declared contrary by laboratory examination using the ZN staining sputum method. Meanwhile, the inclusion criterion of positive-TB samples was when a positively suspected TB patient was declared positive-TB using ZN staining sputum method. Exclusion criteria from positive-TB samples were +HIV subject and positive-TB in extrapulmonary.

Sensing and sampling processes shown at the first block of Figure 1 are elaborated. The sputum samples of this research were obtained three times. First time, on the first day, the suspected patient's sputum was obtained when the TB-suspected patient visited the Public Health Center for the first time, and then the TB-suspected patient brought a sputum pot containing the sample on the second day, early in the morning. The second sample was the sputum saved at home, soon after the TB-suspected patient woke up in the early morning on the second day. On that day, the TB-suspected patient brought the sample and handed it over in the sputum pot to the medical officer in the Public Health Center. Third, the sputum sample was accepted again in Public Health Center on the second day, when the TB-suspected patient gave the morning sputum sample. In a biosafety cabinet, the three sputum samples were combined to meet the minimum volume requirement so that evaporation and sensing processes could be carried

out. The sputum evaporation of a sample was performed by warming it up at the temperature between 38 and 58°C for 60 min in the electronic-nose sample room. The sensors of electronic-nose device during the process used VOC as the biomarker to detect the presence of MTB. Then, the sensors calculated the rate of VOC and converted the chemistry magnitude to electrical magnitude (analog signal). After that, an interface application in the electronic-nose device transformed the analog signal of its sensor to a digital signal. The digital signal was presented in the form of response curve from each sensor with its maximum length of 20 cycles. Subsequently, each cycle of the response curve from the electronic-nose device was read or sampled every 5 sec to obtain digital data.

At the second block of Figure 1, the processes consist of two types of pre-processing, namely, noise reduction and feature extraction. Considering that the response curve of the electronic-nose came out from the first block still contained noises, a noise reduction method became essential to apply. The noises could be reduced using a filter as can be seen in Equation (4), and its results were saved in a memory.

Generally, a finite impulse response (FIR) filter as in Smith (1999) for the case of N-Points is shown in the following equation:

$$y[n] = \sum_{k=0}^M b_k x[n-k] \quad (2)$$

Equation (2) can be written as in Kusumoputro et al. (1999) as:

$$y[n] = \frac{1}{N} \sum_{k=0}^{N-1} x[n-k] \quad (3)$$

and as in Smith (1999), the case of three-point occurs when $M = 2$ and $b_0 = b_1 = b_2 = 1/3$ so that Equation (2) would be the Moving Average Filter Equation used in this research:

$$y[n] = \frac{1}{3} (x[n-1] + x[n] + x[n+1]) \quad (4)$$

The second process of the second block (pre-processing) in Figure 1 is feature extraction (the feature of positive-TB and feature of negative-TB) from each curve of the electronic-nose response. The feature extraction could be done after the noises significantly decrease. If marker events in this research could be recognized in each curve of the electronic-nose response, the feature extraction in this research (as pre-processing) could be performed through searching peaks and troughs as landmarks or features (James,

2007). These searching cycles were used to obtain relative amplitudes (A_r) from each cycle in the electronic-nose response curve. A_r became the features and primary data in this research. Response curves of the three sensors of electronic nose were not involved in the processes of feature extraction as an observation unit, because each sensor had the unique response when sensing was performed so that the calculation of A_r was performed for each response curve individually. The response signal of each sensor illustrated amplitude in time function forming a pattern (Kusumoputro et al., 2002). In this research, features of the sensor response signal were extracted using the following equation, as shown in a study Hardoyo et al. (2015) and Figure 2:

$$A_r = \frac{A_{\max} - A_0}{A_0} \quad (5)$$

Here, A_0 is the sensor response calculated by the time of 'odorant IN' and A_{\max} is the sensor response calculated by the time of 'odorant OUT.'

At the third block, there are three processes in the hybrid genetic algorithm-fuzzy learning vector quantization (HGA-FLVQ) model as displayed in Figure 1. In the first process, HGA read and normalized data from the pre-processing result (all of the training data) of each sensor's output in the pre-processing block.

The second process was the process of assisting initialization of FLVQ artificial neural network so that FLVQ could have an optimal initial cluster center. This process executed HGA operators in the HGA-FLVQ model. Then, the HGA method took the first two A_r s of each class (positive-TB and negative-TB) from the training data to become two parents. After that the two parents were processed by simulated-annealing recombination (SAR) function and simulated-

annealing mutation (SAM) function in HGA (Adler, 1993). A crossover operator in SAR function and a mutation operator in SAM function were implemented repeatedly. Every time HGA produced two offsprings, the values of fitness from these offsprings were also calculated to determine whether those two offsprings were accepted in the new generation population. If the offspring fitness value was better than the parent fitness value, then that offspring was used to replace its parent. However, if not, then it was ignored, and the A_r cluster center of the final generation of this HGA was used for initialization of FLVQ artificial neural network.

The third process of HGA-FLVQ model was model training performed by FLVQ artificial neural network. In this research, this model training was carried out repeatedly using the sputum samples of a group of 98 patients consisting of 49 positive ZN staining patients, 40 negative ZN staining patients, and 9 healthy volunteers who had never been diagnosed with TB by a doctor. The training process using the FLVQ method can be elaborated as follows. Initialization of the initial cluster center was performed by taking HGA output of each sensor. Then, some variables were calculated repeatedly until the desired minimum error was obtained. Those variables were the weighting exponent (also referred to as fuzziness parameter), learning-rate factor (α), learning-rate, new cluster center, and error. Afterward, the last cluster center of each sensor data of each class was saved in a memory as a result of training.

This section also elaborates the result of HGA-FLVQ model training as displayed in Table 2. The process of this HGA-FLVQ model training from TGS813, TGS822, and TGS2611 sensors lasted for 5 to 8 iterations since the error decreased below 0.0001. The error reduction process occurred on each iteration as shown in Figures 3,4,5.

Training processes using LVQ and FLVQ methods as comparator methods are elaborated in the following parts. Learning vector quantization (LVQ) aiming at classifying training data of n vectors into c clusters is one of the learning methods in the artificial neural network (Kusumadewi dan and Hartati, 2010). LVQ is a simple way algorithm applied to the multiclass problem and the complexity of LVQ can be controlled during training phase according to specific needs (Elly et al., 2013).

In the training process of the LVQ method, the weight vector from the input unit to the output unit was referred to as a reference vector (codebook) for the class that the output unit represents. The output unit was positioned by adjusting the weight through supervised training. In this case, the target (a set of




Figure 2: A Cycle of Gas Sensor (TGS) Response in an E-nose Device.

Table 2. Results of HGA-FLVQ model training.

Iteration	Fuzziness parameter	Learning rate	Cluster center	Error
<i>TGS813 Sensor data</i>				
1	1.100590	0.000987	0.001400	0.108494
2	1.101180	0.001541	0.000924	0.140709
3	1.101770	0.002089	0.000798	0.163612
4	1.102360	0.002508	0.000750	0.177564
5	1.102950	0.002808	0.000727	0.186139
<i>TGS822 Sensor data</i>				
1	1.100590	0.001144	0.001169	0.093718
2	1.101180	0.001008	0.001355	0.097408
3	1.101770	0.000919	0.001562	0.102777
4	1.102360	0.000857	0.001782	0.107080
5	1.102950	0.000807	0.002042	0.110949
6	1.103540	0.000769	0.002332	0.114258
7	1.104130	0.000742	0.002620	0.116887
8	1.104720	0.000723	0.002887	0.118887
<i>TGS2611 Sensor data</i>				
1	1.100590	0.000644	0.005906	0.031570
2	1.101180	0.001916	0.000857	0.055587
3	1.101770	0.002333	0.000767	0.081313
4	1.102360	0.003203	0.000704	0.096225
5	1.102950	0.003725	0.000683	0.103321

patterns with known classification) was provided along with the initial distribution of the reference vector. The training using the LVQ method classified the input vector into the same class as the output unit that has a weight vector closest to the input vector (Widodo, 2005).

In this research, the result of the training using LVQ method is presented in Table 3. The stability of weight value appeared after the 10th iteration, either for the negative or positive class used the training data of each sensor. It could be seen that the weight value either of the negative or positive




Figure 3: Error Value Decrease in HGA-FLVQ Model Training using TGS813 Sensor Data.




Figure 4: Error Value Decrease in HGA-FLVQ Model Training using TGS822 Sensor Data.




Figure 5: Error Value Decrease in HGA-FLVQ Model Training using TGS2611 Sensor Data.

class for the last iterations was stable at a certain number.

On the other hand, the training process using the FLVQ method is elaborated as follows. At first, an initialization of the initial cluster center was performed by taking the first item of relative amplitude of each sensor. Then, the next process calculated some variables repeatedly until the desired minimum error was obtained. The variables used were weighting exponent (also referred to as a fuzziness parameter),

Table 3. Results of weight calculation and revision of LVQ training.

	TGS813 Sensor	TGS822 Sensor	TGS2611 Sensor
<i>1st Iteration</i>			
The weight of the negative class	-0.051673025	0.26946884	-0.003179715
The weight of the positive class	-0.002036426	0.1985224	0.21423542
<i>2nd Iteration</i>			
The weight of the negative class	-0.0356808	0.28372976	-3.8830974E-4
The weight of the positive class	0.028204879	0.14191468	0.16531087
<i>3rd Iteration</i>			
The weight of the negative class	-0.0029952978	0.3213401	0.00271485
The weight of the positive class	0.052159406	0.10385808	0.13945574
<i>4th Iteration</i>			
The weight of the negative class	0.03716613	0.37145796	0.0097912615
The weight of the positive class	0.062625006	0.088705875	0.1361857
<i>5th Iteration</i>			
The weight of the negative class	0.07502196	0.407575	0.017326174
The weight of the positive class	0.06537599	0.080875896	0.12186478
<i>6th Iteration</i>			
The weight of the negative class	0.09876694	0.42662817	0.02240251
The weight of the positive class	0.06590489	0.07604013	0.1046309
<i>7th Iteration</i>			
The weight of the negative class	0.11628124	0.44089812	0.026141549
The weight of the positive class	0.06780671	0.074578255	0.09289357
<i>8th Iteration</i>			
The weight of the negative class	0.12786691	0.44939327	0.02856244
The weight of the positive class	0.06953817	0.07365509	0.08497784
<i>9th Iteration</i>			
The weight of the negative class	0.13485843	0.45396066	0.029982805
The weight of the positive class	0.07066574	0.07268719	0.08011716
<i>10th Iteration</i>			
The weight of the negative class	0.13828559	0.4564701	0.030701837
The weight of the positive class	0.07150341	0.072542615	0.07794089

Table 4. Results of cluster center revisions of FLVQ training.

Iteration	Fuzziness parameter	Learning rate	Learning rate	Cluster-Center	Cluster-Center	Error
<i>TGS813 Sensor data</i>						
1	1.100590	0.001075	0.001401	0.073105	0.072574	0.004639
2	1.101180	0.001309	0.001160	0.079481	0.068403	0.000058
<i>TGS822 Sensor data</i>						
1	1.100590	0.001113	0.001346	0.160899	0.142446	0.004211
2	1.101180	0.001361	0.001012	0.218137	0.103790	0.004771
3	1.101770	0.001516	0.000935	0.236398	0.101749	0.000338
4	1.102360	0.001731	0.000869	0.246578	0.106190	0.000123
5	1.102950	0.001982	0.000817	0.257252	0.110136	0.000129
6	1.103540	0.002266	0.000776	0.268349	0.113576	0.000135
7	1.104130	0.002558	0.000747	0.278854	0.116367	0.000118
8	1.104720	0.002829	0.000726	0.287788	0.118478	0.000084
<i>TGS2611 Sensor data</i>						
1	1.100590	0.003575	0.000705	0.040732	0.029254	0.001988
2	1.101180	0.002175	0.000788	0.074589	0.016031	0.001321
3	1.101770	0.002993	0.000715	0.092969	0.017238	0.000339
4	1.102360	0.003588	0.000688	0.101591	0.018471	0.000076

learning-rate factors (alpha), learning-rates, new cluster centers, and an error.

Afterward, the last new cluster centers of each sensor data of each class were saved in a memory as the results of training. In this research, the results of training using the FLVQ method can be seen in Table 4. It is shown that the error decreased below 0.0001 before the 10th iteration. Its reduction in error is relatively faster compared to the LVQ method that needs up to the 10th iterations to obtain the stable weight.

The last process in this research was testing the HGA-FLVQ model. The testing process of HGA-FLVQ model consisted of four steps, namely, reading the test data, calculating Euclidean distances to determine the classes of each relative amplitude item, deciding the classes of each suspect patient, and calculating specificity and sensitivity of the model.

The first process was reading the test data. This process read A_r data of every cycle in the electronic-nose response curve resulting from the sensing of a sputum sample of each TB-suspected patient.

The second process was determining the class classification from every A_r data item. In general, if

P is denoted as X_1, X_2, \dots, X_n and if Q is denoted as Y_1, Y_2, \dots, Y_n , then $d(P, Q)$ could be calculated as in Johnson and Wincham, (2011) for each sensor using Equation (6). In this research, P_i is a variable, and P_i is training datum A_r , P_i s are data training A_r from a TB-suspected patient. Q_1 s are the cluster centers of the training data resulting from every sensor sensing sputum sample of negative ZN staining patients. Q_2 s are the cluster centers of the training data produced from every sensor sensing sputum samples of positive ZN staining patients. Then, the total number of Euclidian distances between one variable P_i and three variables Q_1 s was compared with the total number of Euclidian distances between one variable P_i and three variables Q_2 s. Then, the value of variable d_1 was calculated as shown in Equation (7), and the value of variable d_2 was calculated as shown in Equation (8). Then, d_1 was compared with d_2 to determine a prediction class at the variable P_i . If d_1 was lower than d_2 , then the prediction class at this A_r was represented with '1' (negative), but if the result was contrary or d_2 was lower than d_1 , then the prediction class at this A_r was represented with '2'

Table 5. Testing result of a positive ZN staining patient using HGA-FLVQ model.

Amplitude order	Target class	Class distance 1	Class distance 2	Prediction class
1	2	0.369	0.424	1
2	2	0.357	0.392	1
3	2	0.285	0.238	2
4	2	0.262	0.177	2
5	2	0.337	0.318	2
6	2	0.317	0.150	2
7	2	0.321	0.076	2
8	2	0.310	0.087	2
9	2	0.307	0.090	2
10	2	0.290	0.107	2
11	2	0.286	0.111	2
12	2	0.305	0.092	2
13	2	0.286	0.111	2
14	2	0.284	0.113	2
15	2	0.2880	0.109	2
16	2	0.292	0.105	2
17	2	0.280	0.117	2
18	2	0.288	0.111	2
19	2	0.289	0.112	2
20	2	0.280	0.117	2

(positive). This calculation was performed for all Ar variables:

$$d(P,Q) = \sqrt{(x_1 - y_1)^2 + (x_2 - y_2)^2 + \dots + (x_p - y_p)^2} \quad (6)$$

$$d1 = d11(P_i, Q11) + d12(P_i, Q12) + d13(P_i, Q13) \quad (7)$$

$$d2 = d21(P_i, Q21) + d22(P_i, Q22) + d23(P_i, Q23) \quad (8)$$

The third process was deciding the classes of each suspect patient by adding up the quantity of the prediction classes ('1' and '2'). Then, a level of confidence of a new negative class was calculated through the number of prediction classes '1' divided by the total number of the data Ar of the electronic-nose response curve resulting from the sputum sample of each TB-suspected patient. Following this process was the calculation of the level of confidence

of the new positive class, the number of prediction classes '2' divided by the total number of the data Ar of the electronic-nose response curve resulted from the sputum sample of each TB-suspected patient.

If the level of confidence in the new negative class is higher than 50%, then the prediction class for the sample of the TB-suspected patient is negative. On the contrary, if the level of confidence in the new negative class is below 50%, then the prediction class for the sample of the TB-suspected patient is positive. If the level of confidence in the new positive class is higher than 50%, then the prediction class for the sample of the TB-suspected patient is positive. On the other hand, if the level of confidence in the new positive class is below 50%, then the prediction class for the sample of the TB-suspected patient is negative.

The fourth process was calculating specificity and sensitivity of HGA-FLVQ model using comparator methods. Specificity is the number of negatives

Table 6. Confusion matrix of HGA-FLVQ model using ZN staining method.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN staining: No	TN = 22	FP = 4	26
ZN staining: Yes	FN = 1	TP = 23	24
	23	27	

prediction classes as same as the targets (the target classes of test data), and the number is divided by the quantity of all TB-suspected patients. Then, its result is multiplied by 100%. Sensitivity is the number of positives prediction classes as same as the targets, and the number is divided by the quantity of all TB-suspected patients. Then, its result was multiplied by 100%.

Results and discussion

This section explains the result of HGA-FLVQ model testing and the results of the comparator methods testing (LVQ method and FLVQ method). These models and method testings used data from sensing and sampling result of the electronic-nose response curve on the vapor of 50 TB-suspected patients' sputum. Table 5 presents the recapitulation of HGA-FLVQ model testing result of a patient. In Table 5, target class having value '2' denotes positive ZN staining, target class having value '1' denotes negative ZN Staining, prediction class having value '1' denotes negative HGA-FLVQ, and prediction class having value '2' denotes positive HGA-FLVQ.

The performance of the HGA-FLVQ model was measured using a confusion matrix (Smith, 1999) as displayed in Table 6. Referring to Table 6, TN denotes true negative, FN denotes false negative, TP denotes true positive, and FP denotes false positive. Meanwhile, 'Yes' means the presence of bacteria causing TB and 'No' means the absence of bacteria causing TB.

The HGA-FLVQ model yielded 50 predictions from 50 TB-suspected patients who were tested. Among 50 cases analyzed using the ZN staining sputum method, it turned out that 24 patients were diagnosed as positive TB and 26 patients as negative TB. Based on the confusion matrix as presented in Table 6, some performance rates (Kuhn and Johnson, 2013) of the HGA-FLVQ model could be calculated as displayed in Table 7.

Based on the testing results displayed in Table 6, there were five patients with negative ZN staining who were diagnosed as positive TB according to HGA-FLVQ model analysis. These five patients went through a re-examination by ZN staining, and the results showed that four of them were positive, whereas one was negative as displayed in Table 8. Therefore, the confusion matrix in Table 6 has been revised as shown in Table 9. Similarly, the performance rates of HGA-FLVQ model have been edited as shown in the third column in Table 14.

Afterward, a confusion matrix is adapted from Kuhn and Johnson (2013) and Maysam and Mahdi (2016). In this research, the confusion matrixes were made based on Tables 10 and 12 to observe LVQ and FLVQ method performance. Referring to Tables 10 and 12, TN denotes true negative, FN denotes false negative, TP denotes true positive, and FP denotes false positive. Meanwhile, 'Yes' means the presence of bacteria causing TB and 'No' means the absence of bacteria causing TB.

Table 7. Performance rates of HGA-FLVQ model.

Performance rate	Formula	Result
Accuracy	$(TP+TN)/n$	$(23+22)/50 \times 100\% = 90.00\%$
Error rate	$(FP+FN)/n$	$(4+1)/50 \times 100\% = 10.00\%$
Sensitivity (true positive rate)	TP/ZN Staining Yes	$23/24 \times 100\% = 95.83\%$
False positive rate	FP/ZN Staining No	$4/26 \times 100\% = 15.38\%$
Specificity (true negative rate)	TN/ZN Staining No	$22/26 \times 100\% = 84.62\%$
Precision	TP/Predictive Yes	$23/27 \times 100\% = 85.19\%$
Prevalence	TP/n	$23/50 \times 100\% = 46.00\%$

Table 8. Confusion matrix of HGA-FLVQ model using ZN staining re-examination.

<i>n</i> = 4	Re-Examination ZN Staining:	Re-Examination ZN Staining:	
	No	Yes	
ZN Staining: No	TN=1	FP=3	4
ZN Staining: Yes	FN=0	TP=0	0
	1	3	

Table 9. Final confusion matrix of HGA-FLVQ model after a re-examination by ZN staining.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN Staining: No	TN=22	FP=1	23
ZN Staining: Yes	FN=1	TP=26	27
	23	27	

Table 10. Confusion matrix of LVQ method using ZN staining as Gold standard.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN Staining: No	TN=25	FP=1	26
ZN Staining: Yes	FN=3	TP=21	24
	28	22	

Each of LVQ and FLVQ methods has 50 prediction cases, because there were 50 sputum samples of TB-suspected patients tested. The testing results of 50 prediction cases using LVQ method are 22 positive-TB patients and 28 negative-TB patients, as presented in Table 10. Meanwhile, the testing results of 50 prediction cases using the FLVQ method are 18 positive-TB patients and 32 negative-TB patients, as

Table 11. Final confusion matrix of LVQ method after a re-examination by ZN staining.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN Staining: No	TN=22	FP=4	26
ZN Staining: Yes	FN=3	TP=21	24
	25	25	

Table 12. Confusion matrix of FLVQ method using ZN staining as Gold standard.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN Staining: No	TN=25	FP=1	26
ZN Staining: Yes	FN=7	TP=17	24
	32	18	

Table 13. Final confusion matrix of FLVQ method after a re-examination by ZN staining.

<i>n</i> = 50	Predicted:	Predicted:	
	No	Yes	
ZN Staining: No	TN=22	FP=4	26
ZN Staining: Yes	FN=7	TP=17	24
	29	21	

shown in Table 12. In comparison, the results of ZN staining method testing are 24 positive ZN staining patients and 26 negative ZN staining patients.

From Table 10, it can be understood that the LVQ method correctly predicted 25 out of 26 negative-TB ZN staining, only one of them is false. The LVQ method correctly predicted 21 out of 24 positive-TB by ZN staining and three of them are false. After a

Table 14. Final results of HGA-FLVQ model performance.

Performance rates	Formulation	Performance of HGA-FLVQ model	Performance of LVQ method	Performance of FLVQ method
Accuracy	(TP+TN)/n	$(26+22)/50 \times 100\% = 96.00\%$	$(21+25)/50 \times 100\% = 92.00\%$	$(17+25)/50 \times 100\% = 84.00\%$
Error rate	(FP+FN)/n	$(1+1)/50 \times 100\% = 4.00\%$	$(1+3)/50 \times 100\% = 8.00\%$	$(1+7)/40 \times 100\% = 16.00\%$
Sensitivity (true positive rate)	TP/(ZN Staining Yes)	$26/27 \times 100\% = 96.30\%$	$21/24 \times 100\% = 87.50\%$	$17/24 \times 100\% = 70.83\%$
False positive rate	FP/(ZN Staining No)	$1/23 \times 100\% = 4.35\%$	$1/26 \times 100\% = 3.85\%$	$1/26 \times 100\% = 3.85\%$
Specificity (true negative rate)	TN/(ZN Staining No)	$22/23 \times 100\% = 95.65\%$	$25/26 \times 100\% = 96.15\%$	$25/26 \times 100\% = 96.15\%$
Precision	TP/Predictive Yes	$26/27 \times 100\% = 96.30\%$	$21/22 \times 100\% = 95.45\%$	$17/18 \times 100\% = 94.44\%$
Prevalence	TP/n	$26/50 \times 100\% = 52.00\%$	$21/50 \times 100\% = 42.00\%$	$17/50 \times 100\% = 34.00\%$

re-examining for four patients resulting negative-TB by ZN staining but positive-TB by the HGA-FLVQ's model, Table 10 has been revised in Table 11. Table 11 edits that the LVQ method correctly predicted 22 out of 26 negative-TB by ZN staining.

Table 12 shows that the FLVQ method correctly predicted 25 out of 26 negative-TB ZN staining, only one of them is false. The FLVQ method accurately predicts 17 out of 24 positive-TB by ZN staining but 7 of those are false. After a re-examining for four patients resulting negative-TB by ZN staining but positive-TB by HGA-FLVQ's model, Table 12 has been revised in Table 13. Table 13 edits that the LVQ method correctly predicts 22 out of 26 negative-TB by ZN staining.

Table 14 lists some variables presented as the percentage that was calculated based on the confusion matrix as shown in Tables 11 and 13. It is utilized to observe the performances of the HGA-FLVQ model using both comparator methods (LVQ and FLVQ methods).

Conclusion

The conclusions of this research are as follows:

1. The result of HGA-FLVQ model developed in this research is that the model has 96.30% sensitivity. The sensitivity of HGA-FLVQ model is better than sensitivities of two compar-

ator methods; the sensitivity of FLVQ method is 70.83% and of LVQ method is 87.50%. Thus, this developed model is recommended for strengthening TB laboratory examination in Public Health Centers in Indonesia and for complementing sputum ZN staining method that has been commonly used.

2. The relative amplitudes in the curve of the electronic-nose sensing results can be used by the HGA-FLVQ model to determine the presence of *Mycobacterium tuberculosis* in each TB-suspected patient's sputum.

Acknowledgments

The authors would like to thank the dissertation advisor Lecturers Mr. Agus Harjoko, PhD and Dr. Azhari for their continual support who have guided the authors throughout the study for many years in Gadjah Mada University. The authors also extend their gratitude to Mr. Kuwat Triyana, D. Eng who lent the authors an electronic-nose, namely, Gama1 to be used throughout this research.

Literature Cited

Adler, D. 1993. Genetic algorithms and simulated annealing: a marriage proposal. *IEEE*, 1104–1109.

- Bruins, M., Rahim, Z., Bos, A., Van De Sande, W.W.J., Endtz, H.Ph., and Van Belkum, A. 2013. Diagnostic of active tuberculosis by e-nose analysis exhaled air. *Tuberculosis* 93(2): 232–8.
- Charibaldi dan, N., and Harjoko, A. 2013. Telaah Pustaka Ciri dan Metode – metode Mycobacterium Tuberculosis. *Indonesia Journal of Electronics and Instrumentation Systems* 3(1): 47–60.
- Dahlan, M.S. 2010. Besar Sampel dan Cara Pengambilan Sampel dalam Penelitian Kedokteran dan Kesehatan. Edisi ke 3, Salemba Medika, Jakarta.
- Elly, M.I., Jatmiko, W., and Tjan, B. 2013. Electrocardiogram for biometrics by using Adaptive Multilayer Generalized Learning Vector Quantization (AMGLVQ): integrating feature extraction and classification. *International Journal on Smart Sensing and Intelligent Systems* 6(5): 1891–1917.
- Fend, R., Kolk, A.H.J., Bessant, C., Buijtel, P., Klatser, P.R., and Woodman, A.C. 2006. Prospects for clinical application of electronic-nose technology to early detection of mycobacterium tuberculosis in culture and Sputum. *Journal of Clinical Microbiology* 44(6): 2039–45.
- Gibson, T., Kolk, A., Reither, K., Kuipers, S., Hallam, V., Chandler, R., Dutta, R., Maboko, L., Jung, J., and Klatser, P. 2009. Predictive detection of tuberculosis using electronic nose technology, CP1137, olfaction and electronic nose. Proceedings of the 13th International Symposium, April 15–17, Brescia: 473–4.
- Hardoyono, F., Triyana, K., and Iswanto, B.H. 2015. Rapid discrimination of Indonesian herbal medicines by using electronic nose base on array of commercial gas sensors. *Applied Mechanics and Materials* 771: 209–12.
- James, G.M. 2007. Curve alignment by moments. *The Annals of Applied Statistics* 1(2): 480–501.
- Johnson, R.A., and Winchurn, D.W. 2011. Applied Multivariate Statistical Analysis, Prentice-Hall International Inc., 1998, in: Febriyana, “Analisis Kluster K-Means dan K-Median pada Indikator Kemiskinan”, Skripsi, UIN Syarif Hidayatullah, Jakarta.
- Kolk, A.H.J., Hoelscher, M., Maboko, L., Jung, J., Kuiper, S., Cauchi, M., Bessant, C., Van Beers, S., Dutta, R., Gibson, T., and Reither, K. 2010. Electronic-nose technology using sputum samples in diagnosis of patients with tuberculosis. *Journal of Clinical Microbiology* 48(11): 4235–8.
- Kuhn, M., and Johnson, K. 2013. *Applied Predictive Modeling*, Springer-Verlag, New York, NY, USA.
- Kusumadewi dan, S., and Hartati, S. 2010. *Neuro-Fuzzy Integrasi Sistem Fuzzy dan Jaringan Syaraf*, 2nd ed., Graha Ilmu, Yogyakarta.
- Kusumoputro, B., Widyanto, MR., Fanany, MI., and Budiarto, H. 1999. Improvement of artificial odor discriminant system using fuzzy-LVQ neural network. Proceeding of Third International Conference on Computational Intelligence and Multimedia Applications IC-CIMA, September 23–26, New Delhi: 474–8.
- Kusumoputro, B., and Jatmiko, W. 2002. Recognition of odor mixture using fuzzy-LVQ neural networks with matrix similarity analysis. Proceeding of Asia Pacific Conference on Circuits and System APCCAS, October 28–31, Denpasar: 57–62.
- Kusumoputro, B., Jatmiko, W., and Krisnadh, AA. 2002. Sistem Penciuman Elektronik Menggunakan Algoritma FLVQ dan Analisa Matriks Similaritas untuk Mengenal Aroma Campuran. *Prosiding Ilmu Komputer dan Teknologi Informasi* 3(1): 246–50.
- Maysam, J., and Mahdi, M. 2016. Comparison of predictive models for the early diagnosis of diabetes. *Healthcare Informatics Research*, April 22(2): 95–100.
- Pavlou, A.K., Magan, N., Jones, J.M., Brown, J., Klatser, P., and Turner, A.P.F. 2004. Detection of Mycobacterium Tuberculosis (TB) in vitro and in situ using an electronic nose in combination with a neural network system. *Biosensor and Bioelectronics* 20(3): 538–44.
- Phillips, M., Basa-Dalay, V., Bothamley, G., Caltaneo, R.N., Lam, P.K., Natividad, M.P.R., Schmitt, P., and Wai, J. 2010. Breath biomarkers of active pulmonary tuberculosis. *Tuberculosis* 90(2): 145–54.
- Polikar, R., Shinar, R., Honavar, V., Udpa, L., and Porter, M.D. 2001. Detection and identification of odorants using an electronic nose. *IEEE* 5: 3137–40.
- Smith, S.W. 1999. *The Scientist and Engineer's Guide to Digital Signal Processing*, 2nd ed., California Technical Publishing, San Diego, CA.
- Wang, Y., and Liu, L. 2014. New intelligent classification method based on improved MEB algorithm. *International Journal on Smart Sensing and Intelligent Systems* 7(1): 72–95.
- Widodo, T.S. 2005. *Sistem Neuro Fuzzy untuk Pengolahan Informasi, Pemodelan, dan Kendali*, 1st ed., Graha Ilmu, Yogyakarta.
- Zhou, L., He, X., He, D., Wang, K., and Qin, D. 2011. Biosensing technologies for mycobacterium tuberculosis detection: status and new development. *Clinical and Developmental Immunology* 2011(1): 1–8.